



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005**

	<p><b>Parameter(s) of Accreditation</b>            Dimensional            Electromagnetics – DC/Low Frequency            Time and Frequency            Mechanical            Electromagnetics – RF/Microwave            Thermodynamic</p> <p>This laboratory is compliant to ANSI/NCSL Z540-1-1994; Part 1. ( 20/A01)</p>
--	---

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <sup>Notes 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Uncertainty ( <i>k</i> =2) <sup>Note 3</sup>	Remarks
<b>DIMENSIONAL</b>			
<b>ANGULAR (20/D01)</b>			
Angles	0° to 360° 0° to 360°	0.0022° 0.009°	CMM OGP Vision System
<b>LENGTH &amp; DIAMETER; STEP GAGES (20/D05)</b>			
Micrometers – Outside, Inside, Depth Field calibrations Available <sup>Note 4</sup>	0.01 in to 1 in 1 in to 2 in 2 in to 6 in 6 in to 12 in 12 in to 16 in 16 in to 24 in  24 in to 80 in  24 in to 80 in	22 μin + 1 μin/in 6 μin + 16 μin/in 32 μin + 3 μin/in 41 μin + 3 μin/in 9 μin + 7 μin/in 85 μin + 2.3 μin/in  610 μin  130 μin	Comparison to Gage Blocks       End Standards and CMM; Zero Setting  CMM; 0 in to 1 in Micrometer Travel
Anvil Flatness Field calibrations Available <sup>Note 4</sup>	0 in to 1 in	6.5 μin	Optical Flats

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <sup>Notes 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) <sup>Note 3</sup>	Remarks
Calipers – Outside, Inside, Depth, Step Field calibrations Available <sup>Note 4</sup>	0 in to 1 in 1 in to 2 in 2 in to 6 in 6 in to 12 in 12 in to 16 in 16 in to 24 in 24 in to 80 in	22 $\mu\text{in} + 1 \mu\text{in}/\text{in}$ 6 $\mu\text{in} + 16 \mu\text{in}/\text{in}$ 32 $\mu\text{in} + 3 \mu\text{in}/\text{in}$ 41 $\mu\text{in} + 3 \mu\text{in}/\text{in}$ 9 $\mu\text{in} + 7 \mu\text{in}/\text{in}$ 85 $\mu\text{in} + 2.3 \mu\text{in}/\text{in}$ 640 $\mu\text{in}$	Comparison to Gage Blocks  CMM
Dial Indicators	0 in to 1 in 1 in to 4 in 4 in to 6 in	10 $\mu\text{in} + 19 \mu\text{in}/\text{in}$ 28 $\mu\text{in} + 2 \mu\text{in}/\text{in}$ 46 $\mu\text{in} + 3 \mu\text{in}/\text{in}$	Comparison to Gage Blocks
Length Measurement – Single Axis – Outside	0 in to 1 in 1 in to 10 in 10 in to 13 in 13 in to 80 in	3.2 $\mu\text{in} + 1.0 \mu\text{in}/\text{in}$ 1.6 $\mu\text{in} + 2.5 \mu\text{in}/\text{in}$ 2.8 $\mu\text{in}/\text{in}$ 12 $\mu\text{in} + 1.5 \mu\text{in}/\text{in}$	Pratt & Whitney Labmaster  Laser Interferometer w/CMM
Length Measurement – Single Axis – Inside	0.04 in to 0.5 in 0.5 in to 4 in 4 in to 11 in 11 in to 40 in 40 in to 80 in	5.0 $\mu\text{in} + 4.0 \mu\text{in}/\text{in}$ 7.0 $\mu\text{in} + 2.0 \mu\text{in}/\text{in}$ 3.0 $\mu\text{in} + 3.0 \mu\text{in}/\text{in}$ 39 $\mu\text{in} + 1.4 \mu\text{in}/\text{in}$ 250 $\mu\text{in}$	Pratt & Whitney Labmaster  Gage Blocks with CMM CMM
Height Gages	0 in to 24 in	85 $\mu\text{in} + 1.0 \mu\text{in}/\text{in}$	Comparison to Gage Blocks
Linear Displacement Field calibrations Available <sup>Note 4</sup>	0 in to 1440 in	1.1 $\mu\text{in} + 0.6 \mu\text{in}/\text{in}$	Laser Interferometer
<b>MEASURING WIRES (20/D07)</b>			
Thread Wires	2 TPI to 120 TPI	9.7 $\mu\text{in}$	Pratt & Whitney Labmaster
<b>SPHERICAL DIAMETER; PLUG/RING GAGES (20/D11)</b>			

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <sup>Notes 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) <sup>Note 3</sup>	Remarks					
Outer Diameters - Plugs	0.001 in to 2 in 2 in to 4 in 4 in to 6 in	6.0 $\mu\text{in} + 1.3 \mu\text{in/in}$ 5.0 $\mu\text{in} + 2.0 \mu\text{in/in}$ 2.5 $\mu\text{in} + 2.6 \mu\text{in/in}$	Pratt & Whitney Labmaster					
<b>SURVEYING RODS and TAPES (20/D13)</b>								
Steel Rules, Tape Measures, and PI Tapes	0 in to 72 in 72 in to 1200 in	490 $\mu\text{in} + 3.5 \mu\text{in/in}$ 130 $\mu\text{in} + 7.3 \mu\text{in/in}$	CMM with Vision					
<b>THREADED PLUGS &amp; RING GAGES (20/D14)</b>								
Threaded Plugs – Outer Pitch Diameter	0 in to 6 in	79 $\mu\text{in} + 1.0 \mu\text{in/in}$	Comparator with Thread Wires					
<b>TWO DIMENSIONAL GAGES (20/D15)</b>								
Linear Dimension – Two Axis (X-Y)	27 in x 27 in	150 $\mu\text{in} + 21 \mu\text{in/in}$	OGP Vision System					
<b>COORDINATE MEASURING MACHINES (20/D16)</b>								
Volumetric	0 in <sup>3</sup> to 1 in <sup>3</sup> 1 in <sup>3</sup> to 216 in <sup>3</sup> 0.125 ft <sup>3</sup> to 18 ft <sup>3</sup> 18 ft <sup>3</sup> to 66 ft <sup>3</sup>	53 $\mu\text{in}$ 120 $\mu\text{in}$ 310 $\mu\text{in}$ 0.0014 in	CMM					
Volumetric Field calibrations Available <sup>Note 4</sup>	37 ft <sup>3</sup>	0.012 in	Articulated Arm					
<b>ELECTROMAGNETICS – DC/LOW FREQUENCY</b>								
Measured Parameter or Device Calibrated	Remarks							
<b>AC to DC VOLTAGE/CURRENT CONVERTERS (20/E01)</b>								
Flatness Relative to 1 kHz Field calibrations Available <sup>Note 4</sup>								
<b>Uncertainty (<math>k=2</math>) in % <sup>Note 3,5</sup></b>								
<b>Frequency Range</b>	Range							
	7 V	2.2 V	700 mV	220 mV	70 mV	22 mV	7 mV	2.2 mV
10 Hz	0.006	0.01	0.01	0.015	0.017	0.019	0.02	0.022

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

Measured Parameter or Device Calibrated	Remarks							
<b>AC to DC VOLTAGE/CURRENT CONVERTERS (20/E01)</b>								
Flatness Relative to 1 kHz								
Field calibrations Available <sup>Note 4</sup>								
<b>Uncertainty (k=2) in %</b> <sup>Note 3,5</sup>								
<b>Frequency Range</b>	<b>Range</b>							
	7 V	2.2 V	700 mV	220 mV	70 mV	22 mV	7 mV	2.2 mV
20 Hz	0.007	0.011	0.011	0.015	0.017	0.019	0.02	0.022
50 Hz	0.005	0.006	0.006	0.007	0.008	0.008	0.009	0.009
105 Hz	0.007	0.007	0.007	0.008	0.009	0.009	0.010	0.010
200 Hz	0.005	0.006	0.006	0.007	0.008	0.008	0.009	0.009
2 kHz	0.005	0.006	0.006	0.007	0.008	0.008	0.009	0.009
10 kHz	0.006	0.006	0.006	0.007	0.008	0.008	0.009	0.009
20 kHz	0.005	0.006	0.006	0.007	0.008	0.008	0.009	0.009
50 kHz	0.006	0.006	0.006	0.007	0.008	0.008	0.009	0.009
100 kHz	0.005	0.006	0.006	0.007	0.008	0.008	0.009	0.009
200 kHz	0.006	0.006	0.006	0.007	0.008	0.008	0.009	0.009
500 kHz	0.005	0.006	0.006	0.007	0.008	0.008	0.009	0.009
700 kHz	0.011	0.012	0.012	0.013	0.014	0.014	0.015	0.016
1 MHz	0.011	0.012	0.013	0.014	0.014	0.015	0.015	0.016
1.2 MHz	0.012	0.013	0.014	0.014	0.014	0.015	0.016	0.016
2 MHz	0.014	0.015	0.016	0.016	0.016	0.017	0.018	0.018
3 MHz	0.017	0.019	0.02	0.021	0.022	0.023	0.023	0.025
4 MHz	0.019	0.02	0.022	0.022	0.023	0.024	0.024	0.025
6 MHz	0.023	0.025	0.027	0.028	0.028	0.03	0.031	0.032
8 MHz	0.028	0.029	0.031	0.032	0.032	0.034	0.034	0.036
9 MHz	0.027	0.029	0.03	0.031	0.032	0.033	0.034	0.035
10 MHz	0.026	0.029	0.031	0.030	0.031	0.032	0.033	0.035
12 MHz	0.036	0.039	0.041	0.042	0.043	0.045	0.046	0.048
15 MHz	0.037	0.04	0.042	0.042	0.043	0.046	0.047	0.049
17 MHz	0.042	0.045	0.047	0.047	0.048	0.050	0.050	0.052
20 MHz	0.045	0.048	0.051	0.052	0.054	0.057	0.059	0.061
23 MHz	0.068	0.070	0.073	0.074	0.076	0.079	0.081	0.084
26 MHz	0.083	0.084	0.087	0.09	0.093	0.096	0.098	0.10

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

Measured Parameter or Device Calibrated	Remarks							
<b>AC to DC VOLTAGE/CURRENT CONVERTERS (20/E01)</b>								
Flatness Relative to 1 kHz								
Field calibrations Available <sup>Note 4</sup>								
<b>Uncertainty (k=2) in %</b> <sup>Note 3,5</sup>								
<b>Frequency Range</b>	Range							
	7 V	2.2 V	700 mV	220 mV	70 mV	22 mV	7 mV	2.2 mV
28 MHz	0.094	0.096	0.10	0.10	0.11	0.11	0.12	0.12
30 MHz	0.11	0.11	0.11	0.11	0.12	0.12	0.12	0.13

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC)** <sup>Notes 1,2</sup>

Measured Parameter or Device Calibrated	Remarks								
<b>ELECTROMAGNETICS – DC/LOW FREQUENCY</b>									
<b>AC RESISTORS and CURRENT (20/E02)</b>									
AC Current	Fluke A40 Shunts with 792A								
<b>Range</b>	<b>Uncertainty (k=2) in <math>\mu</math>A/A</b> <sup>Note 3</sup>								
	Frequency Range								
	10 Hz	20 Hz	40 Hz	400 Hz	1 kHz	5 kHz	10 kHz	20 kHz	50 kHz
100 $\mu$ A	61	66	56	60	59	72	81	100	130
200 $\mu$ A	62	43	38	36	36	63	73	98	130
1 mA	45	41	31	38	37	37	37	42	46
2 mA	46	31	31	28	28	28	28	31	39
10 mA	110	53	33	31	31	31	30	32	55
20 mA	130	54	33	31	31	31	32	34	63
100 mA	130	53	33	31	31	31	31	40	73
200 mA	120	57	33	31	31	31	31	36	73
1A	110	57	37	35	36	35	38	50	180
2A	110	60	37	35	36	41	38	52	120
10A	160	85	64	63	63	63	63	83	150
20A	230	150	120	110	120	120	140	150	260
50A	176	86	66	66	65	72	76		
100A	168	81	72	62	62	85			

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

Scope Revised: 2014-04-16

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <sup>Notes 1,2</sup>

Measured Parameter or Device Calibrated	Range	Frequency Range	Uncertainty (k=2) <sup>Note 3,5</sup>	Remarks
AC Current – Measuring Equipment Field calibrations Available <sup>Note 4</sup>	0 µA to 220 µA	10 Hz to 20 Hz	0.03 % + 16 nA	Fluke 5720A
		20 Hz to 40 Hz	0.02 % + 10 nA	
		40 Hz to 1 kHz	0.015 % + 8 nA	
		1 kHz to 5 kHz	0.03 % + 12 nA	
		5 kHz to 10 kHz	0.11 % + 65 nA	
	0.22 mA to 2.2 mA	10 Hz to 20 Hz	0.03 % + 40 nA	
		20 Hz to 40 Hz	0.018 % + 35 nA	
		40 Hz to 1 kHz	0.014 % + 35 nA	
		1 kHz to 5 kHz	0.02 % + 110 nA	
		5 kHz to 10 kHz	0.11 % + 650 nA	
	2.2 mA to 22 mA	10 Hz to 20 Hz	0.039 % + 400 nA	
		20 Hz to 40 Hz	0.018 % + 350 nA	
		40 Hz to 1 kHz	0.014 % + 350 nA	
		1 kHz to 5 kHz	0.02 % + 550 nA	
5 kHz to 10 kHz		0.11 % + 5 µA		
22 mA to 220 mA	10 Hz to 20 Hz	0.033 % + 4 µA		
	20 Hz to 40 Hz	0.018 % + 3.5 µA		
	40 Hz to 1 kHz	0.014 % + 2.5 µA		
	1 kHz to 5 kHz	0.021 % + 3.5 µA		
	5 kHz to 10 kHz	0.11 % + 10 µA		
0.22 A to 2.2 A	20 Hz to 1 kHz	0.027 % + 35 µA		
	1 kHz to 5 kHz	0.046 % + 80 µA		
	5 kHz to 10 kHz	0.7 % + 160 µA		
2.2 A to 11 A	40 Hz to 1 kHz	0.05 % + 170 µA	Fluke 5720A+5725A	
	1 kHz to 5 kHz	0.10 % + 380 µA		
	5 kHz to 10 kHz	0.36 % + 750 µA		
11 A to 20 A	10 Hz to 100 Hz	0.097 % + 2.0 mA	Fluke 5520A	

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

Scope Revised: 2014-04-16

### CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <sup>Notes 1,2</sup>

Measured Parameter or Device Calibrated	Range	Frequency Range	Uncertainty (k=2) <sup>Note 3,5</sup>	Remarks
Extended Frequency Ranges	20 A to 100 A	100 Hz to 1 kHz 1 kHz to 5 kHz 10 Hz to 1 kHz	0.12 % + 5.0 mA 2.3 % + 5.0 mA 0.12 %	Ohms Lab Shunt with Current Source
	29 µA to 330 µA 0.33 mA to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA	10 kHz to 30 kHz	1.2 % + 0.4 µA 0.78 % + 0.6 µA 0.31 % + 4.0 µA 0.31 % + 0.20 mA	Fluke 5520A
	Clamp-on Ammeter Toroidal Type Field calibrations Available <sup>Note 4</sup>	20 A to 150 A	45 Hz to 65 Hz	0.30 % + 26 mA
20 A to 150 A 150 A to 1000 A 150 A to 1000 A		65 Hz to 440 Hz 45 Hz to 65 Hz 65 Hz to 440 Hz	0.83 % + 47 mA 0.35 % + 0.12 A 1.1 % + 0.22 A	Fluke 5520A
Clamp-on Ammeter Non-Toroidal Type Field calibrations Available <sup>Note 4</sup>	20 A to 150 A	45 Hz to 65 Hz	0.57 % + 0.25 A	
	20 A to 150 A 150 A to 1000 A 150 A to 1000 A	65 Hz to 440 Hz 45 Hz to 65 Hz 65 Hz to 440 Hz	1.0 % + 0.25 A 0.60 % + 0.9 A 1.3 % + 0.92 A	
AC Current – Measure Field calibrations Available <sup>Note 4</sup>	0 µA to 100 µA	10 Hz to 20 Hz	0.46 % + 30 nA	Agilent 3458A opt 2
		20 Hz to 45 Hz 45 Hz to 5 kHz	0.18 % + 30 nA 0.072 % + 30 nA	
	0.1 mA to 1 mA	10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz	0.46 % + 0.20 µA 0.17 % + 0.20 µA 0.071 % + 0.20 µA 0.037 % + 0.20 µA	
		1 mA to 10 mA	10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz	

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <sup>Notes 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Frequency Range	Uncertainty (k=2) <sup>Note 3,5</sup>	Remarks
	10 mA to 100 mA	10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz	0.46 % + 20 μA 0.17 % + 20 μA 0.071 % + 20 μA 0.037 % + 20 μA	
	0.1 A to 1 A	10 Hz to 20 Hz 20 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 5 kHz	0.46 % + 0.20 mA 0.19 % + 0.20 mA 0.097 % + 0.20 mA 0.12 % + 0.20 mA	
	1 A to 5 A	3 Hz to 10 kHz	0.19 % + 1.6 mA	Fluke 8846
Impedance Field calibrations Available <sup>Note 4</sup>	5 A to 10 A	3 Hz to 10 kHz	0.16 % + 3 mA	Fluke 8846
	10 A to 100 A	10 Hz to 1 kHz	0.12 %	Ohm Labs Shunt
	1 Ω	10 Hz to 1 MHz	0.059 %	Quadtech 7600 LCR Meter
	10 Ω		0.059 %	
	100 Ω		0.059 %	
	1 kΩ 10 kΩ 100 kΩ		0.059 % 0.062 % 0.19 %	

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <sup>Notes 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Uncertainty (k=2) <sup>Note 3,5</sup>	Remarks
<b>DC RESISTANCE AND CURRENT (20/E05)</b>			
DC Resistance – Measuring Equipment and Measure	0 Ω (floor) 100 μΩ to < 1 mΩ 1 mΩ to < 10 mΩ 10 mΩ to < 100 mΩ 0.1 Ω to < 1 Ω	0.63 nΩ 6.3 μΩ/Ω 3.0 μΩ/Ω 2.3 μΩ/Ω 2.3 μΩ/Ω	Standard Resistors with Current Source and DMM

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology





# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

Scope Revised: 2014-04-16

### CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <sup>Notes 1,2</sup>

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) <sup>Note 3,5</sup>	Remarks
	1 $\Omega$ to < 10 $\Omega$	0.18 $\mu\Omega/\Omega$	Standard Resistors with MI 6242B Bridge
	10 $\Omega$ to < 100 $\Omega$	0.29 $\mu\Omega/\Omega$	
	100 $\Omega$ to < 1 k $\Omega$	0.28 $\mu\Omega/\Omega$	
	1 k $\Omega$ to < 10 k $\Omega$	0.44 $\mu\Omega/\Omega$	
	10 k $\Omega$ to < 100 k $\Omega$	0.19 $\mu\Omega/\Omega$	
	100 k $\Omega$ to < 1 M $\Omega$	1.7 $\mu\Omega/\Omega$	
	1 M $\Omega$ to < 10 M $\Omega$	5.7 $\mu\Omega/\Omega$	
	10 M $\Omega$ to < 100 M $\Omega$	11 $\mu\Omega/\Omega$	Standard Resistor
	100 M $\Omega$ to < 1 G $\Omega$	29 $\mu\Omega/\Omega$	Standard Resistors with 8508 (Double Substitution)
	1 G $\Omega$ to < 20 G $\Omega$	140 $\mu\Omega/\Omega$	

DC Resistance – Measure	100 M $\Omega$ Source	24 $\mu\Omega/\Omega$	Standard Resistor
	1 G $\Omega$ Source	38 $\mu\Omega/\Omega$	
Resistance Ratio	1 $\Omega$ to 1 k $\Omega$	0.2 $\mu\Omega/\Omega$	MI 6242B Bridge
DC Resistance – Measuring Equipment and Measure Field calibrations Available <sup>Note 4</sup>	10 m $\Omega$ to 10 $\Omega$	18 $\mu\Omega/\Omega$ + 50 $\mu\Omega$	HP3458A with Decade Resistor
	10 $\Omega$ to 100 $\Omega$	15 $\mu\Omega/\Omega$ + 500 $\mu\Omega$	
	100 $\Omega$ to 1 k $\Omega$	12 $\mu\Omega/\Omega$ + 500 $\mu\Omega$	
	1 k $\Omega$ to 10 k $\Omega$	12 $\mu\Omega/\Omega$ + 5 m $\Omega$	
	10 k $\Omega$ to 100 k $\Omega$	12 $\mu\Omega/\Omega$ + 50 m $\Omega$	
	100 k $\Omega$ to 1 M $\Omega$	19 $\mu\Omega/\Omega$ + 2 $\Omega$	
	1 M $\Omega$ to 10 M $\Omega$	62 $\mu\Omega/\Omega$ + 100 $\Omega$	
DC Resistance – Measuring Equipment Field calibrations Available <sup>Note 4</sup>	10 M $\Omega$ to 100 M $\Omega$	590 $\mu\Omega/\Omega$ + 1 k $\Omega$	IET HRRS-B-7-100k-5kV
	100 M $\Omega$ to 1 G $\Omega$	0.58 % + 10 k $\Omega$	
	10 M $\Omega$ to 100 M $\Omega$	0.08 %	
	100 M $\Omega$ to 1 G $\Omega$	0.25 %	
	1 G $\Omega$ to 10 G $\Omega$	0.41 %	
	10 G $\Omega$ to 100 G $\Omega$	0.84 %	
	100 G $\Omega$ to 1 T $\Omega$	2.5 %	

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <sup>Notes 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) <sup>Note 3,5</sup>	Remarks
DC Current – Measuring Equipment and Measure	0 A (floor)	0.40 nA	Standard Shunts with Current Source
	100 nA to < 1 $\mu$ A	17 $\mu$ A/A	
	1 $\mu$ A to < 10 $\mu$ A	8.3 $\mu$ A/A	
	10 $\mu$ A to < 100 $\mu$ A	7.9 $\mu$ A/A	
	100 $\mu$ A to < 1 mA	4.5 $\mu$ A/A	
	1 mA to < 10 mA	2.0 $\mu$ A/A	
	10 mA to < 100 mA	1.9 $\mu$ A/A	
	100 mA to < 1 A	2.2 $\mu$ A/A	
	1 A to < 10 A	3.3 $\mu$ A/A	
	10 A to < 20 A	6.9 $\mu$ A/A	
	20 A	61 $\mu$ A/A	
	> 20 A to 100 A	78 $\mu$ A/A	

DC Current – Measuring Equipment and Measure Field calibrations Available <sup>Note 4</sup>	0 $\mu$ A to 100 $\mu$ A	26 $\mu$ A/A + 0.8 nA	HP3458A with Current Source
	100 $\mu$ A to 1 mA	26 $\mu$ A/A + 5.0 nA	
	1 mA to 10 mA	26 $\mu$ A/A + 50 nA	
	10 mA to 100 mA	43 $\mu$ A/A + 0.5 $\mu$ A	
	100 mA to 1 A	0.13 mA/A + 10 $\mu$ A	
Clamp-on Ammeter Non-Toroidal Type Field calibrations Available <sup>Note 4</sup>	1 A to 20 A	44 $\mu$ A/A + 1.8 mA	Shunt w/3458A and current source
	20 A to 100 A	120 $\mu$ A/A	
	20 A to 150 A	0.51 % + 0.14 A	
	150 A to 1000 A	0.51 % + 0.50 A	5520A w/5500A Coil

**DC VOLTAGE (20/E06)**

DC Voltage – Fixed Points Measuring Equipment and Measure	0 V (floor)	59 nV	Ratio Metric with Zener Reference
	1 mV	120 $\mu$ V/V	
	10 mV	22 $\mu$ V/V	
	100 mV	0.80 $\mu$ V/V	
	1 V	0.29 $\mu$ V/V	

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) <sup>Note 3,5</sup>	Remarks
DC Voltage – Measuring Equipment and Measure Field calibrations Available <sup>Note 4</sup>	10 V	0.20 $\mu\text{V/V}$	HP3458A opt 2 with 5720A
	19 V	0.48 $\mu\text{V/V}$	
	100 V	0.28 $\mu\text{V/V}$	
	1000 V	0.54 $\mu\text{V/V}$	
	20 mV to < 50 mV	13 $\mu\text{V/V} + 0.059 \mu\text{V}$	
	50 mV to < 100 mV	7.3 $\mu\text{V/V} + 0.059 \mu\text{V}$	
	100 mV to $\leq 1 \text{ kV}$	0.61 $\mu\text{V/V} + 0.059 \mu\text{V}$	
	0 V to 100 mV	7.1 $\mu\text{V/V} + 0.5 \mu\text{V}$	
	100 mV to 10 V	5 $\mu\text{V/V} + 0.5 \mu\text{V}$	
	10 V to 100 V	7.6 $\mu\text{V/V} + 30 \mu\text{V}$	
100 V to 500 V	11 $\mu\text{V/V} + 100 \mu\text{V}$		
500 V to 800 V	16 $\mu\text{V/V} + 100 \mu\text{V}$		
800 V to 1100 V	21 $\mu\text{V/V} + 100 \mu\text{V}$		
	> 1.1 kV to 5 kV	110 mV/kV	High Voltage Divider Comparison
	> 5 kV to 120 kV	96 mV/kV	

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Remarks												
<b>LOW FREQUENCY AC VOLTAGE (20/E09)</b>													
AC Voltage – Measure	Fluke 792A												
	<b>Uncertainty (<math>k=2</math>) in <math>\mu\text{V/V}</math> <sup>Note 3</sup></b>												
	Frequency Range												
<b>Range</b>	10 Hz	20 Hz	40 Hz	100 Hz	1 kHz	10 kHz	20 kHz	50 kHz	100 kHz	300 kHz	500 kHz	800 kHz	1 MHz
2 mV	630	890	930	920	370	840	840	430	1000	1200	2100	2400	2500
6 mV	330	310	330	210	270	260	280	340	530	760	930	610	660
10 mV	140	190	110	240	190	190	170	220	290	240	650	390	490
20 mV	82	73	63	120	73	84	62	150	210	300	460	350	400
60 mV	100	66	52	45	37	47	52	70	140	270	350	410	430
100 mV	38	35	22	32	30	34	31	41	76	160	200	280	210

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Remarks												
<b>LOW FREQUENCY AC VOLTAGE (20/E09)</b>													
AC Voltage – Measure	Fluke 792A												
	<b>Uncertainty (k=2) in <math>\mu\text{V/V}</math> Note 3</b>												
	Frequency Range												
<b>Range</b>	10 Hz	20 Hz	40 Hz	100 Hz	1 kHz	10 kHz	20 kHz	50 kHz	100 kHz	300 kHz	500 kHz	800 kHz	1 MHz
200 mV	42	24	38	23	19	25	18	45	87	130	120	260	210
600 mV	30	39	17	11	12	11	12	22	31	96	90	86	110
1 V	86	35	18	9.2	9.4	11	7.7	18	10	70	95	95	75
2 V	25	20	18	8	6.9	6.5	7.1	15	11	96	95	76	76
6 V	33	21	18	8.9	6.8	8.2	7	21	12	92	93	78	78
10 V	20	32	18	10	8.3	8.3	9.3	10	11	70	99	110	130
20 V	37	21	19	8.9	8.5	8.9	10	11	15	85	91	78	82
60 V	37	23	19	11	11	12	11	33	16	73			
100 V	86	35	19	15	14	15	11	31	16				
200 V	44	24	19	18	12	12	13	17	21				
600 V	96	47	27	23	16	16	17	46	51				
1000 V	54	21	22	22	19	18	20	49	52				

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Range	Frequency Range	Uncertainty (k=2) Note 3,5	Remarks
AC Voltage – Measure Field calibrations Available Note 4	0 mV to 10 mV	1 Hz to 40 Hz	0.039 % + 3 $\mu\text{V}$	Agilent 3458A opt 2
		40 Hz to 1 kHz	0.028 % + 1 $\mu\text{V}$	
		1 kHz to 20 kHz	0.038 % + 1 $\mu\text{V}$	
		20 kHz to 50 kHz	0.15 % + 1 $\mu\text{V}$	
		50 kHz to 100 kHz	0.59 % + 1 $\mu\text{V}$	
		100 kHz to 300 kHz	4.6 % + 2 $\mu\text{V}$	
	10 mV to 100 mV	1 Hz to 40 Hz	0.013 % + 4 $\mu\text{V}$	
		40 Hz to 1 kHz	0.0094 % + 2 $\mu\text{V}$	
		1 kHz to 20 kHz	0.017 % + 2 $\mu\text{V}$	

*Walter R. M. [Signature]*

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Range	Frequency Range	Uncertainty ( $k=2$ ) <sup>Note 3,5</sup>	Remarks
AC Voltage – Measuring Equipment and Measure	100 mV to 1 V	20 kHz to 50 kHz	0.037 % + 2 $\mu$ V	
		50 kHz to 100 kHz	0.093 % + 2 $\mu$ V	
		100 kHz to 300 kHz	0.36 % + 10 $\mu$ V	
		300 kHz to 1 MHz	1.2 % + 10 $\mu$ V	
	1 V to 10 V	1 Hz to 40 Hz	0.0098 % + 40 $\mu$ V	
		40 Hz to 1 kHz	0.0094 % + 20 $\mu$ V	
		1 kHz to 20 kHz	0.017 % + 20 $\mu$ V	
		20 kHz to 50 kHz	0.036 % + 20 $\mu$ V	
		50 kHz to 100 kHz	0.093 % + 20 $\mu$ V	
		100 kHz to 300 kHz	0.35 % + 100 $\mu$ V	
		300 kHz to 1 MHz	1.2 % + 100 $\mu$ V	
	10 V to 100 V	1 Hz to 40 Hz	0.0095 % + 0.4 mV	
		40 Hz to 1 kHz	0.0095 % + 0.2 mV	
		1 kHz to 20 kHz	0.017 % + 0.2 mV	
		20 kHz to 50 kHz	0.036 % + 0.2 mV	
		50 kHz to 100 kHz	0.093 % + 0.2 mV	
		100 kHz to 300 kHz	0.35 % + 1 mV	
		300 kHz to 1 MHz	1.2 % + 1 mV	
	100 V to 700 V	1 Hz to 40 Hz	0.024 % + 4 mV	
		40 Hz to 1 kHz	0.024 % + 2 mV	
1 kHz to 20 kHz		0.024 % + 2 mV		
20 kHz to 50 kHz		0.041 % + 2 mV		
50 kHz to 100 kHz		0.14 % + 2 mV		
700 V to 2 kV	100 kHz to 300 kHz	0.46 % + 10 mV		
	300 kHz to 1 MHz	1.7 % + 10 mV		
	1 Hz to 40 Hz	0.047 % + 40 mV		
	40 Hz to 1 kHz	0.047 % + 20 mV		
	1 kHz to 20 kHz	0.071 % + 20 mV		
		20 kHz to 50 kHz	0.14 % + 20 mV	
		50 kHz to 100 kHz	0.35 % + 20 mV	
		60 Hz	0.05 % + 1.6 V	High Voltage Divider Comparison

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Range	Frequency Range	Uncertainty ( $k=2$ ) <sup>Note 3,5</sup>	Remarks
AC Voltage – Measuring Equipment Field calibrations Available <sup>Note 4</sup>	2 kV to 10 kV	60 Hz	0.11 % + 0.4 V	Fluke 5720A
	10 kV to 80 kV	60 Hz	0.12 %	
	0 mV to 2.2 mV	10 Hz to 20 Hz	0.16 % + 4 μV	
		20 Hz to 40 Hz	0.1 % + 4 μV	
		40 Hz to 20 kHz	0.077% + 4 μV	
		20 kHz to 50 kHz	0.12 % + 4.5 μV	
		50 kHz to 100 kHz	0.17 % + 5 μV	
		100 kHz to 300 kHz	0.33 % + 10 μV	
		300 kHz to 500 kHz	0.47 % + 20 μV	
		500 kHz to 1 MHz	0.58 % + 20 μV	
	2.2 mV to 22 mV	10 Hz to 20 Hz	0.044 % + 4 μV	
		20 Hz to 40 Hz	0.031 % + 4 μV	
		40 Hz to 20 kHz	0.015 % + 4 μV	
		20 kHz to 50 kHz	0.031 % + 4 μV	
		50 kHz to 100 kHz	0.059 % + 5 μV	
		100 kHz to 300 kHz	0.12 % + 12 μV	
		300 kHz to 500 kHz	0.16 % + 20 μV	
		500 kHz to 1 MHz	0.3 % + 20 μV	
	22 mV to 220 mV	10 Hz to 20 Hz	0.028 % + 12 μV	
		20 Hz to 40 Hz	0.011 % + 7 μV	
40 Hz to 20 kHz		0.0083 % + 7 μV		
20 kHz to 50 kHz		0.021 % + 7 μV		
50 kHz to 100 kHz		0.047 % + 17 μV		
100 kHz to 300 kHz		0.092 % + 20 μV		
300 kHz to 500 kHz		0.14 % + 25 μV		
500 kHz to 1 MHz		0.28 % + 45 μV		
220 mV to 2.2 V	10 Hz to 20 Hz	0.028 % + 40 μV		
	20 Hz to 40 Hz	0.01 % + 15 μV		
	40 Hz to 20 kHz	0.0048 % + 8 μV		
	20 kHz to 50 kHz	0.0081 % + 10 μV		

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <sup>Notes 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Frequency Range	Uncertainty ( $k=2$ ) <sup>Note 3,5</sup>	Remarks	
	2.2 V to 22 V	50 kHz to 100 kHz	0.012 % + 30 $\mu$ V		
		100 kHz to 300 kHz	0.043 % + 80 $\mu$ V		
		300 kHz to 500 kHz	0.1 % + 200 $\mu$ V		
		500 kHz to 1 MHz	0.18 % + 300 $\mu$ V		
	22 V to 220 V	10 Hz to 20 Hz	0.028 % + 0.4 mV		
		20 Hz to 40 Hz	0.01 % + 0.15 mV		
		40 Hz to 20 kHz	0.0049 % + 50 $\mu$ V		
		20 kHz to 50 kHz	0.0083 % + 0.1 mV		
		50 kHz to 100 kHz	0.011 % + 0.2 mV		
		100 kHz to 300 kHz	0.03 % + 0.6 mV		
		300 kHz to 500 kHz	0.10 % + 2 mV		
		500 kHz to 1 MHz	0.17 % + 3.2 mV		
	220 V to 1100 V	10 Hz to 20 Hz	0.028 % + 4 mV		Fluke 5720A/5725A
		20 Hz to 40 Hz	0.01 % + 1.5 mV		
		40 Hz to 20 kHz	0.0056 % + 0.6 mV		
		20 kHz to 50 kHz	0.0093 % + 1 mV		
50 kHz to 100 kHz		0.016 % + 2.5 mV			
100 kHz to 300 kHz		0.085 % + 16 mV			
300 kHz to 500 kHz		0.44 % + 40 mV			
220 V to 750 V	40 Hz to 1 kHz	0.011 % + 4 mV			
	1 kHz to 30 kHz	0.017 % + 6 mV			
	220 V to 750 V	30 kHz to 50 kHz	0.061 % + 11 mV		
		50 kHz to 100 kHz	0.23 % + 45 mV		

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <sup>Notes 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) <sup>Note 3</sup>	Remarks
<b>LOW FREQUENCY CAPACITANCE (20/E10)</b>			
Capacitance – Measure: 10 Hz to 1 MHz	1.1 $\mu$ F to 10 mF	0.06 % + 30 aF	Quadtech 7600 LCR Bridge
	0.33 mF	0.048 %	Time/Charge Method with HP3458A
	0.8 mF	0.027 %	

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) <sup>Note 3</sup>	Remarks
	1 mF	0.024 %	
	1.2 mF	0.023 %	
	3 mF	0.018 %	
	3.3 mF	0.017 %	
	8 mF	0.016 %	
	10 mF	0.016 %	
	12 mF	0.016 %	
	30 mF	0.015 %	
	80 mF	0.014 %	
	100 mF	0.014 %	

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Range	Frequency Range	Uncertainty ( $k=2$ ) <sup>Note 3,5</sup>	Remarks
Capacitance – Measuring Equipment Field calibrations Available <sup>Note 4</sup>	0.19 nF to 3.3 nF	10 Hz to 10 kHz	0.39 % + 7.8 pF	Fluke 5520A
	3.3 nF to 11 nF	10 Hz to 1 kHz	0.21 % + 7.8 pF	
	11 nF to 110 nF	10 Hz to 1 kHz	0.21 % + 78 pF	
	110 nF to 330 nF	10 Hz to 1 kHz	0.21 % + 0.23 nF	
	0.33 μF to 1.1 μF	10 Hz to 600 Hz	0.20 % + 0.78 nF	
	1.1 μF to 3.3 μF	10 Hz to 300 Hz	0.20 % + 2.3 nF	
	3.3 μF to 11 μF	10 Hz to 150 Hz	0.20 % + 7.8 nF	
	11 μF to 33 μF	10 Hz to 120 Hz	0.31 % + 23 nF	
	33 μF to 110 μF	10 Hz to 80 Hz	0.35 % + 78 nF	
	110 μF to 330 μF	DC to 50 Hz	0.35 % + 0.23 μF	
	0.33 mF to 1.1 mF	DC to 20 Hz	0.35 % + 0.78 μF	
	1.1 mF to 3.3 mF	DC to 6 Hz	0.35 % + 2.3 μF	
	3.3 mF to 11 mF	DC to 2 Hz	0.35 % + 7.8 μF	
	11 mF to 33 mF	DC to 0.6 Hz	0.58 % + 23 μF	
	33 mF to 110 mF	DC to 0.2 Hz	0.85 % + 78 μF	

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) <sup>Note 3,5</sup>	Remarks
<b>LOW FREQUENCY INDUCTANCE (20/E11)</b>			

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology





**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) <sup>Note 3,5</sup>	Remarks
Inductance – Source Field calibrations Available <sup>Note 4</sup>	100 mH	0.07 %	Fixed Inductors with Quadtech 7600
Inductance – Measure: 1 kHz Field calibrations Available <sup>Note 4</sup>	10 $\mu$ H to 100 H	0.07 %	Quadtech 7600

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Range	Frequency Range	Uncertainty ( $k=2$ ) <sup>Note 3,5</sup>	Remarks
<b>LF POWER/ENERGY (20/E12)</b>				
Power – Measuring Equipment AC Power <sup>Note 8</sup> (PF = 1) 3.3 mA to 9 mA	0.11 mW to 3 mW 3 mW to 9 W	10 Hz to 65 Hz 10 Hz to 65 Hz	0.13 % 0.077 %	Fluke 5520A
9 mA to 33 mA	0.3 mW to 10 mW 10 mW to 33 W	10 Hz to 65 Hz 10 Hz to 65 Hz	0.089 % 0.077 %	
33 mA to 90 mA	1 mW to 30 mW 30 mW to 90 W	10 Hz to 65 Hz 10 Hz to 65 Hz	0.071 % 0.057 %	
90 mA to 330 mA	3.0 mW to 100 mW 100 mW to 300 W	10 Hz to 65 Hz 10 Hz to 65 Hz	0.089 % 0.078 %	
0.33 A to 0.9 A	11 mW to 300 mW 300 mW to 900 W	10 Hz to 65 Hz 10 Hz to 65 Hz	0.071 % 0.081 %	
0.9 A to 2.2 A	30 mW to 720 mW 720 mW to 2 kW	10 Hz to 65 Hz 10 Hz to 65 Hz	0.089 % 0.079 %	
2.2 A to 4.5 A	80 mW to 1.4 W 1.4 W to 4.5 kW	10 Hz to 65 Hz 10 Hz to 65 Hz	0.088 % 0.18 %	
4.5 A to 20.5 A	150 mW to 6.7 W	10 Hz to 65 Hz	0.17 %	
	6.7 W to 20 kW	10 Hz to 65 Hz	0.17 %	

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Range	Frequency Range	Uncertainty (k=2) Note 3,5	Remarks
<b>PHASE METERS (20/E15)</b>				
AC Phase – Measure	0° to 360°	1 Hz to 1 kHz > 1 kHz to 50 kHz > 50 kHz to 200 kHz	0.0020° 0.0041° 0.010°	Clark-Hess 5002 Bridge Set
AC Phase – Generate:  50 mV to 100 V	0° to 360°	1 Hz to 1 kHz > 1 kHz to 6.25 kHz > 6.25 kHz to 50 kHz > 50 kHz to 200 kHz	0.013° 0.017° 0.021° 0.050°	Clark-Hess 5500-2 Phase Standard
100 V to 120 V	0° to 360°	1 Hz to 1 kHz > 1 kHz to 6.25 kHz > 6.25 kHz to 50 kHz > 50 kHz to 200 kHz	0.017° 0.026° 0.037° 0.095°	

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Range	Uncertainty (k=2) Note 3,5	Remarks
<b>TIME AND FREQUENCY</b>			
<b>FREQUENCY DISSEMINATION (20/F01)</b>			
Frequency – Source and Measure Note 9	10 MHz	5.8 x 10 <sup>-10</sup> Hz	Rubidium Frequency Standard
<b>OSCILLATOR CHARACTERISTICS (20/F03)</b>			
Total Harmonic Distortion: CW, Modulation	5 Hz to 500 kHz 500 kHz to 1 MHz	1.4 dB 2.3 dB	Krohn-Hite 6900B
Harmonic Distortion	100 kHz to 2.9 GHz 2.9 GHz to 6.5 GHz 6.5 GHz to 13.2 GHz 13.2 GHz to 22 GHz	1.7 dB 1.9 dB 2.6 dB 2.9 dB	Spectrum Analyzer

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <sup>Notes 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) <sup>Note 3,5</sup>	Remarks
	22 GHz to 26.5 GHz	3.7 dB	
<b>PULSE WAVEFORM (20/F04)</b>			
Rise time (Generate)	≥ 14 ps	2 ps	Pulser
Rise time (Measure)	28 ps to 300 ps 300 ps to 5 ns 5 ns to 100 ns 100 ns	14 % 4.7 % 1.4 % 0.73 %	Sampling System
<b>MECHANICAL</b>			
<b>MASS DETERMINATION (20/M08)</b>			
Mass – Metric Field calibrations Available <sup>Note 4</sup>	8 kg 7 kg 6 kg 5 kg 4 kg 2 kg 1 kg 500 g 200 g 100 g 50 g 20 g 10 g 5 g 2 g 1 g 500 mg 200 mg 100 mg 50 mg 20 mg 10 mg 5 mg 2 mg	18 mg 16 mg 15 mg 13 mg 12 mg 10 mg 1.6 mg 1.0 mg 0.39 mg 0.35 mg 0.11 mg 0.16 mg 0.14 mg 0.13 mg 0.16 mg 0.083 mg 0.078 mg 0.076 mg 0.065 mg 0.064 mg 0.073 mg 0.084 mg 0.097 mg 0.064 mg	Echelon III

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) <sup>Notes 1,2</sup>**

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) <sup>Note 3,5</sup>	Remarks
	1 mg	0.064 mg	
<b>TORQUE (20/M15)</b>			
Torque Measure Field calibrations Available <sup>Note 4</sup>	2 lbf-in to 811 lbf-ft	1.0 %	Torque Calibrator
<b>WEIGHING INSTRUMENTS (20/M16)</b>			
Balance and Scale Calibration Field calibrations Available <sup>Note 4</sup>	114 kg	0.60 g	ASTM Class 1 and Class 2 Mass Standards
	64 kg	0.45 g	
	32 kg	0.21 g	
	16 kg	98 mg	
	8 kg	18 mg	
	7 kg	16 mg	
	6 kg	15 mg	
	5 kg	13 mg	
	4 kg	12 mg	
	2 kg	10 mg	
	1 kg	1.6 mg	
	500 g	1.0 mg	
	200 g	0.39 mg	
	100 g	0.35 mg	
	50 g	0.11 mg	
	20 g	0.16 mg	
	10 g	0.14 mg	
	5 g	0.13 mg	
	2 g	0.16 mg	
	1 g	0.083 mg	
500 mg	0.078 mg		
200 mg	0.076 mg		
100 mg	0.065 mg		
50 mg	0.064 mg		
20 mg	0.073 mg		
10 mg	0.084 mg		
5 mg	0.097 mg		
	2 mg	0.064 mg	

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) Note 3,5	Remarks
	1 mg	0.064 mg	

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2**

Measured Parameter or Device Calibrated	Range	Frequency Range	Uncertainty ( $k=2$ ) Note 3	Remarks
<b>ELECTROMAGNETICS – RF/MICROWAVE</b>				
<b>RF/MICROWAVE POWER (20/R17)</b>				
RF Power Field calibrations Available Note 4	+30 dBm to -20 dBm	100 kHz to 2.6 GHz	0.11 dBm	Measuring Receiver
	-20 dBm to -40 dBm	2.6 GHz to 26.5 GHz	0.16 dBm	Power Meter w/ sensor
Sine Flatness (RF)		9 kHz to 6 GHz	0.18 dBm	Agilent E4418B w/ E9304A
		9 kHz to 6GHz	0.073 dB	
<b>SCATTERING PARAMETERS (20/R18)</b>				
Attenuation	0 dB to -10 dB -10 dB to -20 dB -20 dB to -30 dB -30 dB to -40 dB -40 dB to -50 dB -50 dB to -60 dB -60 dB to -70 dB -70 dB to -80 dB -80 dB to -90 dB -90 dB to -100 dB -100 dB to -110 dB -110 dB to -120 dB	2.5 MHz to 26.5 GHz	0.02 dB 0.021 dB 0.025 dB 0.029 dB 0.034 dB 0.039 dB 0.043 dB 0.052 dB 0.056 dB 0.06 dB 0.12 dB 0.18 dB	Agilent 8902 with 11793A sensor

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC)** <sup>Notes 1,2</sup>

Measured Parameter or Device Calibrated	Range	Uncertainty ( <i>k</i> =2) <sup>Note 3,5</sup>	Remarks	
<b>THERMODYNAMIC</b>				
<b>HUMIDITY (20/T02)</b>				
Humidity – Measure	0 % RH to 90 % RH	2.2 % RH	Vaisala M170 & HMP76	
Field calibrations Available <sup>Note 4</sup>	90 % RH to 100 % RH	2.7 % RH		
<b>LABORATORY THERMOMETERS (20/T03)</b>				
Measuring Equipment	-25 °C to 25 °C	0.020 °C	PRT with Liquid Baths	
	25 °C to 150 °C	0.027 °C		
	150 °C to 200 °C	0.033 °C		
	200 °C to 300 °C	0.040 °C		
	300 °C to 420 °C	0.01 % + 0.44 °C		PRT with Metrology Well
	420 ° to 660 °C	0.021 %		
	0° to 50°C	0.007 °C	Liquid Bath with RTD	
Measure	-200 °C to 0 °C	0.017 °C	PRT and Hart 1560	
Field calibrations Available <sup>Note 4</sup>	0 °C to 420 °C	0.024 °C		
	420 °C to 660 °C	0.031 °C		
<b>PRESSURE (20/T05)</b>				
Absolute Pressure Source – Pneumatic	0.2 psi to 1.45 psi	0.0013 % + 0.29 mpsi	Ruska 2465 DHI FPG 7601	
	1.45 psi to 50 psi	0.0015 % + 14 μpsi		
	50 psi to 1000 psi	0.0019 %		
Gage Pressure Source – Pneumatic	-14.7 psi to -0.2 psi	0.0013 % + 8 μpsi	Ruska 2465 DHI FPG 8601 Ruska 2465	
	-60 inH <sub>2</sub> O to +60 inH <sub>2</sub> O	0.0028 % + 22 μinH <sub>2</sub> O		
	0.2 psi to 100 psi	0.0013 % + 8 μpsi		
	100 psi to 1000 psi	0.0019 %		
Gage Pressure Source – Hydraulic	72.5 psi to 7250 psi	0.0020 % + 2.9 mpsi	DHI PG7000	
	200 psi to 20 000 psi	0.0035 % + 7.3 mpsi		
	725 psi to 43 500 psi	0.0047 % + 15 mpsi		
	43 500 psi to 72 500 psi	0.0055 %		
Determination of Piston Area	0.2 psi to 100 psi	0.0011 %	Ruska 2465	

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



# National Voluntary Laboratory Accreditation Program



## CALIBRATION LABORATORIES

NVLAP LAB CODE 200730-0

Scope Revised: 2014-04-16

### CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) <small>Note 3,5</small>	Remarks
Pressure Source Pneumatic Field calibrations Available <small>Note 4</small>	100 psi to 1000 psi	0.0017 %	DHI PG7000 (Gas to 6 kpsi)
	72.5 psi to 7250 psi	0.0017 %	
	200 psi to 20 000 psi	0.0026 %	
	725 psi to 72 500 psi	0.0051 %	
	Hydraulic Field calibrations Available <small>Note 4</small>	-14.7 psi to -0.2 psi	0.0013% + 8 $\mu$ psi
0.2 psi to 100 psi		0.0013% + 8 $\mu$ psi	Ruska 2465 absolute mode Ruska 2465 gauge or absolute
0.2 psi to 100 psi		0.0011% + 0.29 mpsi	
	100 psi to 1000 psi	0.0019 %	
	50 psi to 15 000 psi	0.011 % + 0.17 psi	Ametek T-150
<b>THERMOCOUPLES &amp; PYROMETER INDICATORS (20/T08)</b>			
Isothermal Block Verification Field calibrations Available <small>Note 4</small>	Ambient (~ 23 °C)	0.04 °C	Thermocouple Half Junction
Electrical Simulation of Thermocouples	Type J	-210 °C to 1200 °C	Thermocouple Half Junction
	Type K	-270 °C to 1372 °C	
	Type T	-270 °C to 400 °C	
	Type E	-270 °C to 1000 °C	Fluke 5520A
	Type R	0 °C to 250 °C	
	250 °C to 400 °C	0.36 °C	
	400 °C to 1000 °C	0.35 °C	
	1000 °C to 1767 °C	0.47 °C	
Type S	0 °C to 250 °C	0.48 °C	
	250 °C to 1000 °C	0.37 °C	
	1000 °C to 1400 °C	0.38 °C	

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary  
Laboratory Accreditation Program**



**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**CALIBRATION AND MEASUREMENT CAPABILITIES (CMC)** Notes 1,2

Measured Parameter or Device Calibrated	Range	Uncertainty ( $k=2$ ) <small>Note 3,5</small>	Remarks
	1400 °C to 1767 °C	0.47 °C	
Type N	-200 °C to -100 °C	0.41 °C	
	-100 °C to -25 °C	0.24 °C	
	-25 °C to 120 °C	0.22 °C	
	120 °C to 410 °C	0.21 °C	
Type B	410 °C to 1300 °C	0.29 °C	
	600 °C to 800 °C	0.45 °C	
	800 °C to 1000 °C	0.35 °C	
	1000 °C to 1550 °C	0.32 °C	
	1550 °C to 1820 °C	0.35 °C	
<b>END</b>			

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology





**CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200730-0**

Scope Revised: 2014-04-16

**Notes**

**Note 1:** A Calibration and Measurement Capability (CMC) is a description of the best result of a calibration or measurement (result with the smallest uncertainty of measurement) that is available to the laboratory's customers under normal conditions, when performing more or less routine calibrations of nearly ideal measurement standards or instruments. The CMC is described in the laboratory's scope of accreditation by: the measurement parameter/device being calibrated, the measurement range, the uncertainty associated with that range (see note 3), and remarks on additional parameters, if applicable.

**Note 2:** Calibration and Measurement Capabilities are traceable to the national measurement standards of the U.S. or to the national measurement standards of other countries and are thus traceable to the internationally accepted representation of the appropriate SI (Système International) unit.

**Note 3:** The uncertainty associated with a measurement in a CMC is an expanded uncertainty using a coverage factor,  $k = 2$ , with a level of confidence of approximately 95 %. Units for the measurand and its uncertainty are to match. Exceptions to this occur when marketplace practice employs mixed units, such as when the artifact to be measured is labeled in non-SI units and the uncertainty is given in SI units (Example: 5 lb weight with uncertainty given in mg).

**Note 3a:** The uncertainty of a specific calibration by the laboratory may be greater than the uncertainty in the CMC due to the condition and behavior of the customer's device and specific circumstances of the calibration. The uncertainties quoted do not include possible effects on the calibrated device of transportation, long term stability, or intended use.

**Note 3b:** As the CMC represents the best measurement results achievable under normal conditions, the accredited calibration laboratory shall not report smaller uncertainty of measurement than that given in a CMC for calibrations or measurements covered by that CMC.

**Note 3c:** As described in Note 1, CMCs cover calibrations and measurements that are available to the laboratory's customers under *normal conditions*. However, the laboratory may have the capability to offer special tests, employing special conditions, which yield calibration or measurement results with lower uncertainties. Such special tests are not covered by the CMCs and are outside the laboratory's scope of accreditation. In this case, NVLAP requirements for the labeling, on calibration reports, of results outside the laboratory's scope of accreditation apply. These requirements are set out in Annex A.1.h. of NIST Handbook 150, Procedures and General Requirements.

**Note 4:** Uncertainties associated with field service calibration may be greater as they incorporate on-site environmental contributions, transportation effects, or other factors that affect the measurements. (This note applies only if marked in the body of the scope.)

**Note 5:** Values listed with percent (%) are percent of reading or generated value unless otherwise noted.

**Note 6:** NVLAP accreditation is the formal recognition of specific calibration capabilities. Neither NVLAP nor NIST guarantee the accuracy of individual calibrations made by accredited laboratories.

**Note 7:** See [NIST Handbook 150](#) for further explanation of these notes.

**Note 8:** The uncertainties shown are for the most favorable conditions. There is an increase in uncertainty that corresponds to the laboratory's AC voltage and current uncertainties at different frequencies other than the ones shown. Power factors (PF) other than the one shown contribute to the power uncertainty. PF is related to the cosine of phase. Therefore, uncertainties track the laboratory's phase uncertainty closely at PF near one, but are magnified heavily as PF approaches zero. The lab may also report reactive power, apparent power, and power factor under this accreditation. If needed, contact laboratory for more information regarding uncertainties at frequency and power factor combinations other than the ones shown.

**Note 9:** Uncertainty values of derivatives of 10 MHz will differ due to resolution, noise and gating errors.

2014-01-01 through 2014-12-31

Effective dates

For the National Institute of Standards and Technology